



tel. +420 545 555 111, fax +420 545 222 728, www.cmi.cz

EC-TYPE EXAMINATION CERTIFICATE

Number: TCM 142/13 - 5009

Addition 2

This addition replaces all previous versions of this certificate in full wording.

Page 1 from 9 pages

In accordance:

with Directive 2004/22/EC of the European Parliament and of the Council as amended

implemented in Czech Republic by Government Order No. 464/2005 Coll. as amended that

lays down technical requirements on measuring instruments.

Manufacturer:

ECOMESS Sp. z o.o.

UI. Szczawińska 42c

95-100 Zgierz

Poland

For:

water meter - single jet

type: Picoflux and i Flux

Accuracy class: 2

Temperature class: T30, T50, T70, T90 or T30/90

Valid until:

27 January 2023

Document No:

0511-CS-A004-13

Description:

Essential characteristics, approved conditions and special conditions, if any, are described in

this certificate.

Date of issue:

29 May 2015

Certificate approved by:

RNDr. Pavel Klenovský